

ANSI/ASB Standard 077, First Edition 2020

**Standard for the Developmental and Internal Validation
of Forensic Serological Methods**



Standard for the Developmental and Internal Validation of Forensic Serological Methods

ASB Approved March 2020

ANSI Approved September 2020



410 North 21st Street
Colorado Springs, CO 80904

This document may be downloaded for free at: www.asbstandardsboard.org

This document is provided by the AAFS Standards Board for free. You are permitted to print and download the document and extracts from the document for your own use, provided that:

- *you do not modify this document or its related graphics in any way;*
- *you do not use any illustrations or any graphics separately from any accompanying text; and,*
- *you include an acknowledgement alongside the copied material noting the AAFS Standards Board as the copyright holder and publisher.*

You expressly agree not to reproduce, duplicate, copy, sell, resell, or exploit for any commercial purposes, this document or any portion of it. You may create a hyperlink to www.asbstandardsboard.org to allow persons to download their individual, free copy of this document. Your hyperlink must not portray AAFS, the AAFS Standards Board, this document, our agents, associates and affiliates in an offensive manner, or be misleading or false. You may not use our trademarks as part of your link without our written agreement for you to do so.

The AAFS Standards Board retains the sole right to submit this document to any other forum for any purpose.

Certain commercial entities, equipment or materials may be identified in this document to describe a procedure or concept adequately. Such identification is not intended to imply recommendations or endorsement by the AAFS or the AAFS Standards Board, nor is it intended to imply that the entities, materials, or equipment are necessarily the best available for the purpose.

Foreword

This standard provides requirements for the validation of methods that will be used to evaluate body fluids, stains, or residues related to forensic investigations. A validation should include characterization of the test procedure, limitations of the method, and the identification and documentation of influences that may change performance.

In this document the industry standard definition for serology as it relates to forensic science is used instead of the traditional scientific definition.

This document was revised, prepared, and finalized as a standard by the DNA Consensus Body of the AAFS Standards Board. The draft of this standard was developed by the Biological Methods Subcommittee of the Organization of Scientific Area Committees (OSAC) for Forensic Science.

The AAFS Standards Board (ASB) is an ANSI-accredited Standards Developing Organization with the purpose of providing accessible, high quality science-based consensus forensic standards. The ASB is a wholly owned subsidiary of the American Academy of Forensic Sciences (AAFS), established in 2015 and accredited by the American National Standards Institute (ANSI) in 2016. The ASB consists of Consensus Bodies (CB), which are open to all materially interested and affected individuals, companies, and organizations; a Board of Directors; and Staff.

The following applies to all ASB documents:

the term '**shall**' indicates that a provision is mandatory, and can be audited for compliance

the term '**should**' indicates that a provision is not mandatory, but recommended as good practice.

All hyperlinks and web addresses shown in this document are current as of the publication date of this standard.

Keywords: *forensic serology, validation, internal validation, developmental validation, forensic biology*

Table of Contents

1	Scope.....	1
2	Normative References.....	1
3	Terms and Definitions.....	1
4	Validation Requirements.....	3
4.1	General.....	3
4.2	Developmental Validation.....	4
4.3	Internal Validation.....	4
4.4	Validation Documentation.....	5
5	Conformance.....	5
	Annex A (informative) Bibliography.....	6

Standard for the Developmental and Internal Validation of Forensic Serological Methods

1 Scope

This standard provides requirements for developmental and internal validations of forensic serological methods to evaluate body fluids, stains, or residues related to forensic investigations.

This standard does not address validation of forensic DNA analysis procedures.

2 Normative References

The document contains no normative references. See Annex A, Bibliography for other references.

3 Terms and Definitions

3.1 characterization of the test procedure

Comprises the components of the test, the procedure used, limitations of the test, and the methods for detection and analysis including whether the test is presumptive or confirmatory and the type of body fluid, stain, or residue being targeted.

3.2 confirmatory test

A test that is specific for the presence of a body fluid, stain, or residue of interest, and reduces or eliminates false positive results.

3.3 contamination

The unintentional introduction of exogenous materials or substances into a test sample.

3.4 contamination studies

Experiments performed to assess the risk that unintended material may be introduced into a sample from test assay components, instrumentation, the operator, or test procedures.

3.5 controls

Samples of known type, run in parallel with experimental, reference, or evidence samples that are used to demonstrate that a procedure is working correctly.

3.6 control studies

Experiments performed to establish the necessary controls for each procedure, the frequency with which the controls should be performed (e.g., concurrently, daily, before use, etc.) and the performance expectations for each control.

3.7**developmental validation**

The acquisition of test data and determination of conditions and limitations of a new methodology; this generally occurs while the conditions and parameters are being worked out prior to the establishment of a defined assay, procedure or product. Internal validation studies follow developmental validation studies.

3.8**forensic serology**

The detection, characterization, identification, and/or typing of body tissues and fluids, either in native form or as stains or residues left at a crime scene using physical methods (e.g. normal and enhanced lighting), biochemical assays, reactions and/or microscopy.

3.9**internal validation**

The accumulation of test data within the laboratory for developing the laboratory standard operating procedures and determining the limits of the method(s). Internal validation demonstrates that the established protocols for the technical steps of the test and for data interpretation perform as expected in the laboratory.

3.10**interference studies**

Experiments performed to determine substances that inhibit or affect the intensity of the assay signal.

3.11**mixture studies**

Experiments performed to evaluate the performance of the test method when samples containing mixtures of similar or different body fluids and/or cell types are assayed.

3.12**mock casework samples**

Samples of known origin that mimic or simulate a range of casework sample types that may include laboratory created samples or proficiency test samples.

3.13**performance check**

A quality assurance measure to assess the functionality of laboratory instruments, reagents and equipment that affect the accuracy and/or validity of forensic sample analysis.

3.14**population studies**

Experiments performed to determine the effectiveness and utility of the test with representative samples from the population.

3.15**presumptive test**

A screening test which may be positive in the presence of a biological material of interest. Some presumptive tests are sensitive but not specific. A positive result indicates that further testing could be informative.

3.16**repeatability studies**

Experiments performed to verify the results of the assay by the same personnel and/or applicable instrumentation.

3.17**reproducibility studies**

Experiments performed to assess the capability to obtain the same test results when an experiment is repeated between different operators and/or detection instruments.

3.18**robustness studies**

Experiments performed to measure the capability of a procedure to remain unaffected by small but deliberate variations in method parameters and provide an indication of reliability during normal usage.

3.19**sensitivity studies**

Experiments performed to define the lower and upper limits/bounds of an assay to accurately detect an analyte. Experiments include a serial dilution performed during developmental and/or internal validation of methods.

3.20**specificity studies**

Experiments performed to evaluate the ability of the system to provide reliable results for targeted analytes in the presence of cross-reactive substances.

3.21**technical designee**

The designated individual in the laboratory who has technical responsibility.

4 Validation Requirements**4.1 General**

4.1.1 The laboratory shall use validated methods for all presumptive and confirmatory tests. Developmental and internal validation shall precede the implementation of any new methods used for forensic serological analysis.

4.1.2 The laboratory shall use different samples for developmental validation studies and internal validation studies.

4.1.3 The technical designee shall determine the number, types and range of sample types to be used in each of the validation studies to ensure the generation of sufficient data to validate the method.

4.1.4 The internal validation shall not exceed the scope of the conditions tested in the developmental validation. Samples that fall outside the range of conditions used in developmental validation shall require additional developmental validation studies.

4.1.5 Any change to a validated procedure shall be evaluated to determine if analytical results are affected. If a modification affects the analytical result, the procedure shall require an additional validation prior to implementation. If the analytical results are not affected, the modification shall include a comparison of the modified method to the original method. The evaluation of changes to validated procedures shall conform to requirements of this standard and all comparisons, evaluations, modifications, and validations shall be documented.

4.1.6 For laboratory systems that consist of more than one laboratory, validation studies may be shared; however, performance checks shall be conducted and documented at each site.

4.2 Developmental Validation

A developmental validation study shall include the following:

- characterization of the test procedure,
- control studies;
- interference studies,
- population studies,
- reproducibility studies,
- robustness studies,
- sensitivity studies,
- specificity studies,
- mixture studies,
- the analysis of mock casework samples.

4.3 Internal Validation

An internal validation study shall include the following:

- contamination studies;
- control studies;
- repeatability studies;
- reproducibility studies;
- sensitivity studies;
- specificity studies;

- mixture studies;
- the analysis of mock casework samples.

4.4 Validation Documentation

4.4.1 The laboratory shall identify and maintain a list of the scientific literature describing the test, its limitations, and the scientific principles that serve as a foundation.

4.4.2 The laboratory shall have at a minimum a summary of data from any external developmental validation(s).

4.4.3 Developmental validation documentation shall include any limitations of the method.

4.4.4 If a required component is not performed as part of a developmental or internal validation, documentation must be included to warrant the omission.

4.4.5 The technical designee shall document the review and approval of all validations, modifications to procedures or equipment, and performance checks. Other approvals may be necessary according to laboratory policy.

4.4.6 The technical designee shall review and approve any new or revised standard operating procedures as a result of internal validation or procedure modifications. Other approvals may be necessary according to laboratory policy.

4.4.7 The laboratory shall have a policy for retention of the summary and data correlating to developmental validations, internal validations, modification to the procedures, and performance checks.

5 Conformance

In order to demonstrate conformance with this standard, the laboratory shall have the following:

- a) documentation of the developmental validation which can be internal, external, or documented in peer reviewed literature;
- b) documentation of all internal validation studies, data, and outcomes;
- c) documentation of any modification to the procedures, if applicable;
- d) documentation of performance checks, if applicable; and
- e) documentation of approval by the technical designee prior to implementation.

Annex A (informative)

Bibliography

The following bibliography is not intended to be an all-inclusive list, review, or endorsement of literature on this topic. The goal of the bibliography is to provide examples of publications addressed in the standard.

- 1) Scientific Working Group on DNA Analysis Methods (SWGDM): *Guidelines for the Collection and Serological Examination of Biological Evidence*, issue date 01/15/2015.^a

^a Available from: http://media.wix.com/ugd/4344b0_bce915901bb14b9cb36049df6a8441e2.pdf



Academy Standards Board
410 North 21st Street
Colorado Springs, CO 80904

www.asbstandardsboard.org